

CHECK YOUR KIT

Naloxone kit may be missing needles



Prenoxad injectable naloxone kits may be incomplete.

All supplied naloxone is safe to use but reports from abroad suggest a very small number of kits may contain one or no needles, instead of two.

A naloxone kit held up in the light will show two distinct dark rectangles showing that it contains two needles

You can check whether your kit is affected **WITHOUT BREAKING THE SEAL**



The Lot number is on the label on the yellow naloxone box.



What to do

Hold against a light source; you should easily see the two dark patches of the needle packaging.

If you can only see one dark patch or none then return your naloxone kit to any naloxone supplier, tell them that you are returning it as it does not contain two needles, and ask for a new kit.

Lot numbers that MAY be affected

- 0116917
- 0119973
- 0120140
- 0125553
- 0125555
- 0125724
- 0126941
- 0126943
- 0130732
- 0130843
- 0134251
- 0136031
- 0136536
- 0136551
- 0137656
- 0137768
- 0138525
- 0138904
- 0139907
- 0140236
- 0141035
- 0141812

#checkyourkit

 **SDF**
Scottish Drugs
Forum